

REFERENCE: Limited Tender No. OT-20/MNNIT/Purchase Office/CIR/MAGNETRON SPUTTERING /2012-13 Date: 05.12.2012, for procurement of 'MAGNETRON SPUTTERING' at CIR.

(A.) List of Bidders found NON-RESPONSIVE (along with reasons) In Technical Bid (Part-I) Evaluation
तकनीकी बोली (भाग-1) मूल्यांकन में अनुत्तरदायी (कारणों के साथ) पाए गए बोलीदाताओं की सूची

| Sl. No. | Name & Address of Bidder(s) | बोलीदाता(ओं) का नाम व पता |
|---------|---|---------------------------|
| 1. | M/s Mansha Vacuum Equipments Pvt. Ltd., No. B-107, 'C' Main Road, 6 th Cross, 3 rd Stage, Peenya Industrial Estate, Bangalore | |
| 2. | M/s Vacuum Techniques Pvt. Ltd., No. 2/13, Phase-I, Peenya Industrial Area, Bangalore | |
| 3. | M/s Toshiwal, 267, Kilpauk Garden Road, Chennai-600010 | |
| 4. | M/s Harley Instruments, Plot No. 4, Survey No. 47, vPune Satara Road, Pune | |
| 5. | M/s MACWIN India, T-5, Pankaj Central Market I. P. Extn., New Delhi | |

Reasons for being NON-RESPONSIVE

अनुत्तरदायी होने के कारण

| Sl. No. | Reasons as per point in Tender Document निविदा दस्तावेज में बिंदु के अनुसार कारण | | Name of Bidder(s) बोलीदाता(ओं) का नाम |
|---------|---|------------------|---|
| | In tender document | In submitted bid | |
| 1. | Sl. no. 1 in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; Completely stand alone model | Not Mentioned | M/s Mansha Vacuum Equipments Pvt. Ltd., Bangalore |
| | Sl. no. 4 in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; Horizontal sample mounting facility | Not Mentioned | |
| | Sl. no. 5 in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; Substrates size 2" diameter with Uniformity of thin films over 2" dia substrate ±5% OR Better | Not Mentioned | |
| | Sl. no. 7 in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; The instrument should have provision of rotation of sample holder for uniform deposition. | Not Mentioned | |
| | Sl. no. 12 in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; The system should be connected to turbo molecular pump, Rotary pump, and gauges (of reputed make, such as PFEIFFER/ LEYBOLD/VARIEN/ALCATEL or equivalent). | Not Mentioned | |
| | Sl. no. 15 in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; Suitable capacity water chiller | Not Mentioned | |
| | Sl. no. 16 (General Terms & Condition) in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; The same quoted system or equivalent category system should have at least 10 numbers of installations in last five years in India by the vendor. List of such users and copy of their installation report should be enclosed. | Not Mentioned | M/s Vacuum Techniques Pvt. Ltd., Bangalore M/s Toshiwal, Chennai-600010 M/s Harley Instruments, Pune M/s MACWIN India, New Delhi |

(B.) List of Bidders found RESPONSIVE in Technical Bid (Part-I) Evaluation
तकनीकी बोली (भाग-1) मूल्यांकन में उत्तरदायी पाए गए बोलीदाताओं की सूची

| Sl. No. | Name & Address of Bidder(s) | बोलीदाता(ओं) का नाम व पता |
|---------|---|---------------------------|
| 1. | M/s Hind High Vacuum Company Pvt. Ltd., No. 31,32,33,34 & 37, KIADB Industrail Area, Dabaspet, Nelamangala Taluk, Bangalore Rural | |

SCHEDULE of Price Bid (Part-2) opening

मूल्य बोली (भाग-2) खोलने का कार्यक्रम

Date: 08.04.2013 (Monday)

Time: at 14.30 Hrs.

Venue: Institute Conference Room, MNNIT **ALLAHABAD**

आशीष
 21.03.2013
 Faculty-in-charge
 Purchase
 M.N.N.I.T., Allahabad