

REFERENCE: Open Tender No. OT-21/MNNIT/Purchase Office/CIR/Scanning Probe Microscope/2012-13 Date: 05.12.2012, for procurement of 'Scanning Probe Microscope' at CIR.

(A.) List of Bidders found NON-RESPONSIVE (along with reasons) In Technical Bid (Part-I) Evaluation

तकनीकी बोली (भाग-I) मूल्यांकन में अनुत्तरदायी (कारणों के साथ) पाए गए बोलीदाताओं की सूची

Sl. No.	Name & Address of Bidder(s)	बोलीदाता(ओं) का नाम व पता
1.	M/s Sinsil International 1649-A, Ground Floor, Greenfield Colony, Faridabad	
2.	M/s LABINDIA 372, Udyog Vihar, Phas-II, Gurgaon, Haryana	
3.	M/s MACWIN India, T-5, Pankaj Central Market I. P. Extn., New Delhi	
4.	M/s Edge Tech Scientific Pvt. Ltd. E-Block, Vikas Puri, New Delhi	

Reasons for being NON-RESPONSIVE

अनुत्तरदायी होने के कारण

Sl. No.	Reasons as per point in Tender Document निविदा दस्तावेज में बिंदु के अनुसार कारण		Name of Bidder(s) बोलीदाता(ओं) का नाम
	In tender document	In submitted bid	
1.	Sl. no. 2 in Technical Specifications (C) Annexure-A at Chapter-3 (STC) on page no. 18/30;		M/s Sinsil International, Faridabad
	Small scanner Scanning Range : 10µm x10µm and 2 µm Z range OR Smaller Z RMS noise: 0.03 nm OR better	No 10µm × 10 µm × 3 µm Z RMS: 0.046 nm	
	Large scanner Scanning Range : 90 µm x90 µm and 8 µm Z range OR better Z RMS noise: 0.05 nm OR better	No 100µm × 100 µm × 10 µm; RMS: 0.152 nm	
	Sl. no. 5 (E) in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; Controller (E) Multi Lock in amplifiers 3 OR more	No	M/s MACWIN India, New Delhi
	Sl. no. 16 (General Terms & Condition) in Technical Specifications (C) Annexure-B at Chapter-3 (STC) on page no. 18/30; The same quoted system or equivalent category system should have at least 10 numbers of installations in last five years in India by the vendor. List of such users and copy of their installation report should be enclosed.	No	M/s Sinsil International, Faridabad M/s LABINDIA, Haryana M/s MACWIN India, New Delhi M/s Edge Tech Scientific Pvt. Ltd. New Delhi

(B.) List of Bidders found RESPONSIVE in Technical Bid (Part-I) Evaluation

तकनीकी बोली (भाग-I) मूल्यांकन में उत्तरदायी पाए गए बोलीदाताओं की सूची

Sl. No.	Name & Address of Bidder(s)	बोलीदाता(ओं) का नाम व पता
1.	M/s ICON Analytical Equipment Pvt. Ltd. 501, Pragatideep Building, 5 th floor, Plot No. 08, Laxmi Nagar District Centre, Vikas Marg, New Delhi	
2.	M/s Toshniwal Brothers (SR) Pvt. Ltd., No.11, AECS Layout, 4 th Main, 3 rd Cross, Sanjaynagar I Stage, Geddalahalli, Bangalore	

SCHEDULE of Price Bid (Part-2) opening

मूल्य बोली (भाग-2) खोलने का कार्यक्रम

Date: 08.04.2013 (Monday)

Time: at 15.30 Hrs.

Venue: Institute Conference Room, MNNIT *ALLAHABAD*

आशीष
21.03.2013

Faculty-in-charge
Purchase

M.N.N.I.T., Allahabad