

Minutes of Pre-bid Conference

Tender no.: OT-21/MNNIT/Purchase Office/CIR/Scanning Probe Microscope/2012-13

Date: 05.12.2012

For procurement of: Procurement of various lab equipment

Date & time: 12.12.2012 at 15.00 hrs.

Venue: Institute Conference Room, MNNIT

The following participants attended the conference:

Representatives of MNNIT:

- 1) Prof. R. P. Tiwari
- 2) Prof. Vinod Yadava
- 3) Dr. P. R. Paul
- 4) Dr. Naresh Kumar
- 5) Dy. Registrar (Accounts)
- 6) Faculty In-charge (Purchase)

Representatives of Prospective Bidder's:

Sl. No.	Name of Firm	Represented by
1)	M/s Labindia Instruments Pvt. Ltd., 372, Udyog Vihar, Phase-II, Gurgaon-122016	Sri Mohit Gupta
2)	M/s ICON Analytical Equipment Pvt. Ltd., 501, Pragati Deep Bldg., 5th floor, Plot No. 8, Laxmi Nagar, Vikas Marg, New Delhi-110092	Sri Anilesh Pandey
3)	M/s Macwin India, T-5, Pankaj Central Market, IP Extension, Delhi-110092	Sri Santanu Kundu

Opening Remarks

- (i) The Faculty In-charge (Purchase) had conducted the Pre-Bid Conference and at the beginning welcomed to every body attending the Pre-Bid Conference for the Tender no **OT-21/MNNIT/Purchase Office/CIR/Scanning Probe Microscope/2012-13** Date: 05.12.2012 for procurement of various lab equipments.
- (ii) It was explained that purpose of Pre-Bid Conference is to explain the various important provisions of the bidding documents to the prospective bidders and to clarify any queries that the bidders may have in the subject bidding documents.
- (iii) The indenter, Registrar discussed a brief description about the 'Procurement of various lab equipments', as per tender document, before the audience.
- (iv) The members representing the bidders were asked to furnish their queries in written format so that the replies to the same can provided by the purchaser. Replies to the queries are presented in **Table-1**.
- (v) The Faculty-in-charge (Purchase) expressed his profound gratitude to the participants for their active involvement.
- (vi) The meeting ended with a vote of thanks to the chair.



Faculty-in-charge
Purchase
M.N.N.I.T., Allahabad

Table-1

Minutes of Pre-bid Conference

**Tender no.: OT-21/MNNIT/Purchase Office/CIR/Scanning Probe
Microscope/2012-13 Date: 05.12.2012**

Sl. No.	Query From	Gist of the Query	Response	Remarks (If any)
1.	M/s Labindia Instruments Pvt. Ltd., 372, Udyog Vihar, Phase-II, Gurgaon	“Regarding Specifications of automatic cantilever calibration, what is this mean? It should be cantilever calibration only.”	Specifications mentioned in tender is self explanatory. No further explanation/change is possible.	
2.	M/s ICON Analytical Equipment Pvt. Ltd., 501, Pragati Deep Bldg., 5th floor, Plot No. 8, Laxmi Nagar, Vikas Marg, New Delhi	“Large Area scanner spaces is $90\mu\text{m}\times 90\mu\text{m}\times 8\mu\text{m}$ (XYZ). But Bruker Model has the big Scanner spaces $125\mu\text{m}\times 125\mu\text{m}\times 5\mu\text{m}$ (XYZ). Please change Z range 8m to 5m or more.”	Not accepted	


16/12/12

**Faculty-in-charge
Purchase
M.N.N.I.T., Allahabad**