

No. CHEM/54/2014.

ANIL
820
317119
5-30
03-07-14

Faculty In-charge (Purchase)

Sub: Clarification of Queries against your letter no. 781 dated 02.07.14

Ref: Open tender No. 71/MNNIT/Chemistry/Modular open Architecture Spectrofluorometer/2014 dt. 14.06.2014

With reference to the queries raised by firms regarding the above open tender during pre-bid conference held on 02.07.14 at 1.00 PM, the responses are given below.

- 1) Amendment request raised by M/s Specialise Instruments Marketing Company Unit 305, Kailash Industrial Complex, A-wing 3rd Floor, Vikroli (W) Mumbai 400022

Queries about the specification given in the Sl. No. of "Technical Specifications"/ "Bidder's Compliance Statement", Page no. 18 & 19 of the tender document	Details of the specifications mentioned in Page no. 18 & 19 of the tender document	Response of the Indenter
3	Light source...measurements)	No amendment is possible.
4	Excitation monochromator...	No amendment is possible.
5	Emission monochromator...	No amendment is possible.
7	Detector...	Amended.

The amended form of specification 7 is given below:

7.	Detector: - Photomultiplier Tube (PMT) with detection range: 190 to 680 nm. - Temperature cooled PMT / InGaAs Detector with detection range: 600 to 1700 nm.
----	---

T Ghosh. 03/7/14.
(Dr. Tamal Ghosh)
(Indenter)

Forwarded.
1/7/14
3/7/14