

**REFERENCE:** Open Tender No. OT-61/MNNIT/CIR/UV VIS-NIR Spectrophotometer/2014 dated 14.06.2014.

**(A.) List of Bidders found **NON-RESPONSIVE** (along with reasons) In Technical Bid (Part-I) Evaluation**

तकनीकी बोली (भाग-1) मूल्यांकन में अनुत्तरदायी (कारणों के साथ) पाए गए बोलीदाताओं की सूची

Sl. No.	Name & Address of Bidder(s) बोलीदाता(ओं) का नाम व पता
1.	M/s Perkin Elmer (India) Pvt. Ltd., Community Centre, 3rd Phase I, Narina Industrial Area, Near PVR Cinema, New Delhi-110028

**Reasons for being NON-RESPONSIVE**

**अनुत्तरदायी होने के कारण**

Sl. No.	Reasons as per point in Tender Document निविदा दस्तावेज में बिंदु के अनुसार कारण	Name of Bidder(s) बोलीदाता(ओं) का नाम	
	In tender document	In submitted bid	
1.	Noise: 0.00003 Abs or less (1500 nm) Silt width 2nm, RMS value at point no. 6 in Annexure-A 'Technical Specifications' at Chapter-3 (SCC) on page no. (17/32)	At 2A and 1500 nm $\delta$ 0.00100A	M/s Perkin Elmer (India) Pvt. Ltd., New Delhi-110028

**(B.) List of Bidders found **RESPONSIVE** in Technical Bid (Part-I) Evaluation**

तकनीकी बोली (भाग-1) मूल्यांकन में उत्तरदायी पाए गए बोलीदाताओं की सूची

Sl. No.	Name & Address of Bidder(s) बोलीदाता(ओं) का नाम व पता
1.	(1.) M/s Agilent Technologies India Pvt. Ltd., Unit Nos., 105-116, First Floor, Splendor Forum Plot No. 3, District Centre, Jasola, New Delhi-110025
2.	(2.) M/s Toshvin Analytical Pvt. Ltd., 502, Sethi Bhawan, 7, Rajendra Place, New Delhi-110008

**SCHEDULE of Price Bid (Part-2) opening**

**मूल्य बोली (भाग-2) खोलने का कार्यक्रम**

**Date : 19.09.2014 (Friday)**

**Time : at 10.30 hrs. (10.30 AM)**

**Venue : Conference Room (Purchase Office), MNNIT Allahabad**

संकाय प्रभारी-क्रय/Faculty Incharge Purchase  
मोतीलाल नेहरू राष्ट्रीय प्रौद्योगिकी संस्थान इलाहाबाद  
Motilal Nehru National Institute of Technology Allahabad  
इलाहाबाद-211004 /Allahabad-211004  
भारत/India