

REFERENCE: Open Tender No. OT-70/MNNIT/Chemistry/Elemental Analyzer/2014 dated 14.06.2014.

(A.) List of Bidders found **NON-RESPONSIVE (along with reasons) In Technical Bid (Part-I) Evaluation**

तकनीकी बोली (भाग-1) मूल्यांकन में अनुत्तरदायी (कारणों के साथ) पाए गए बोलीदाताओं की सूची

Sl. No.	Name & Address of Bidder(s) बोलीदाता(ओं) का नाम व पता
1.	(1.) M/s Shiva Global Technologies, 129, Functional Industrial Estate, patparganj, New Delhi-110004
2.	(2.) M/s Spechrome, 45-233, Street No. 2, NMDC Colony, East Anandbagh, Malkajgiri, Hyderabad-500047
3.	(3.) M/s SV Instruments Analytical Pvt. Ltd., B-10, 1 st floor, Mayapuri Industrial Area, Phase-I, New Delhi-110064

Reasons for being NON-RESPONSIVE

अनुत्तरदायी होने के कारण

Sl. No.	Reasons as per point in Tender Document निविदा दस्तावेज में बिंदु के अनुसार कारण	Name of Bidder(s) बोलीदाता(ओं) का नाम	
	In tender document	In submitted bid	
1.	The Instruments must also be capable of working with Argon as carrier Gas at point no. 12 in Annexure-A 'Technical Specifications' at Chapter-3 (SCC) on page no. (19/31)	For the offered Instrument possible on request as only factory settable, No field upgrade possible. On page no. 62 of the bid it is mentioned that the quoted EURO EA 3000 use Helium carrier gas exclusively. Also the firm has changed our specification mentioned in this point in tender document to: "The Instrument must also be capable of working with Helium or Argon as carrier gas" while making the compliance sheet.	M/s Shiva Global Technologies, New Delhi
2.	Filled Cylinders (Each of at least 47 ltrs. Water capacity) on in number for each of the following gases: Oxygen, Helium, Argon, all fitted with regulators must be supplied with the instruments	Not Complied for Argon gas	
3.	'Compliance of Technical Specifications for Annexure-A' in Annexure-B at Chapter-3 (SCC) on page no. (19/31)	Not mentioned in quoted bid	M/s Spechrome, Hyderabad
4.	Analysis times CHN ≤ 6 min, CHNS ≤ 8 min, Oxygen ≤ 4 min at point no. 3 in Annexure-A 'Technical Specifications' at Chapter-3 (SCC) on page no. (19/31)	The offered equipment requires CHN=8 min, CHNS=10min and the time required for oxygen not mentioned in the bidder's compliance statement.	M/s SV Instruments Analytical Pvt. Ltd., New Delhi

(B.) List of Bidders found **RESPONSIVE in Technical Bid (Part-I) Evaluation**

तकनीकी बोली (भाग-1) मूल्यांकन में उत्तरदायी पाए गए बोलीदाताओं की सूची

Sl. No.	Name & Address of Bidder(s) बोलीदाता(ओं) का नाम व पता
1.	(1) M/s Perkin Elmer (India) Pvt. Ltd., Community Centre, 3rd Phase I, Narina Industrial Area, Near PVR Cinema, New Delhi-110028

SCHEDULE of Price Bid (Part-2) opening

मूल्य बोली (भाग-2) खोलने का कार्यक्रम

Date : 19.09.2014 (Friday)

Time : at 12.30 hrs. (12.30 PM)

Venue : Conference Room (Purchase Office), MNNIT Allahabad

संकाय प्रचारी-क्रय/Faculty Incharge Purchase
मोतीलाल नेहरू राष्ट्रीय प्रौद्योगिकी संस्थान इलाहाबाद
Motilal Nehru National Institute of Technology Allahabad
इलाहाबाद-211004 /Allahabad-211004
भारत/India