

Minutes of Pre-bid Conference

Tender no.:	TEQIP-II/MNNIT/93 Dated 15.04.2015
For :	LCR Meter
Date & time:	05.05.2015 at 15.00 Hrs.
Venue:	Conference Room, MNNIT (adjacent to Purchase Office)

The following participants attended the conference:

**Representatives of MNNIT:**

- |                                    |  |
|------------------------------------|--|
| 1- Prof. R.K. Srivastava, Head MED | 5- Nodal Officer (Finance), TEQIP-II               |
| 2- Dr. K.N. Pandey, MED            | 6- Nodal Officer (Procurement), TEQIP-II           |
| 3- Dr. D.K. Shukla, MED            | 7- Dr. Samir Saraswati, Nominated Member, TEQIP-II |
| 4- Coordinator TEQIP-II            | 8- Registrar                                       |

**Representatives of Prospective Bidder's:**


Sl. No.	Name of Firm	Represented by
1.	M/s Kamtron Systems Pvt. Ltd., 402 Eros Apartment, 56 Nehru Place, New Delhi-110019	Mr. Mukesh Bhargava

**Opening Remarks**

- (i) The Nodal Officer (Procurement), TEQIP-II had conducted the Pre-Bid Conference and at the beginning welcomed to everybody attending the Pre-Bid Conference for the aforesaid open tender.
- (ii) It was explained that purpose of Pre-Bid Conference is to explain the various important provisions of the bidding documents to the prospective bidders and to clarify any queries that the bidders may have in the subject bidding documents.
- (iii) The indenter discussed a brief description about the equipment, as per the Tender document, before the audience.
- (iv) The members representing the bidders were asked to furnish their queries in written format so that the replies to the same can be provided by the purchaser. Replies to the queries are presented in **Table-1**.
- (v) The Nodal Officer (Procurement), TEQIP-II expressed his profound gratitude to the participants for their active involvement.
- (vi) The meeting ended with a vote of thanks to the chair.

**Table-1**  
**Minutes of Pre-bid Conference**

Sl. No.	Firm	RFP Reference(s) (Section, Page)	Points in tender document	Points of Clarification Required / Query	Resolution
1.	M/s Kamtron Systems Pvt. Ltd., 402 Eros Apartment, 56 Nehru Place, New Delhi-110019	Section – IV Page No. 42	20 Hz to 2MHz with 4-digit resolution in any range · 0.05% or better basic accuracy with superior measurement repeatability at low and high impedance · High – speed measurements : 5.6 ms or better, PC connectivity (LAN , USB and GPI B)	Technical Specification have not mention applicable cables & Text fixture for testing of objects	Standard probes and test fixtures shall be included in the LCR Meter. Firms should supply standard probes and test fixtures for lead components such as: <ul style="list-style-type: none"><li>• 4 terminal probes</li><li>• test fixtures etc.</li></ul>

  
संकाय प्रभारी-क्रय / Faculty Incharge Purchase  
मोतीलाल नेहरू राष्ट्रीय प्रौद्योगिकी संस्थान इलाहाबाद  
Motilal Nehru National Institute of Technology Allahabad  
इलाहाबाद-211004 Allahabad 211004  
भारत/India